



BS-230

Clinical Chemistry Analyzer

Technical Specifications

System function

Automatic, Discrete, Random Access, Bench-top

STAT sample priority

Throughput: Up to 200 tests/hour, up to 400 tests/hour with ISE

Measuring principles: Absorbance photometry, Turbidimetry, Ion

Selective Electrode technology

Methodology: End-point, Fixed-time, Kinetic, optional ISE,

Single/Dual/ reagent chemistries,

monochromatic / bi-chromatic

Original system pack reagent ready to use

Close system and open system is optional

Reagent/Sample Handling

Reagent/Sample tray: 80 positions for reagents and 40 positions

for samples in 24-hour refrigerated

compartment (2~12°C)

Reagent volume: 100~250μl, step by 0.5μl Sample volume: 2~45μl, step by 0.1μl

Reagent/Sample probe: Liquid level detection, vertical collision

protection and inventory checking, reagent pre-warming

Probe cleaning: Automatic washing for interior and exterior

Carry over < 0.05%

Automatic sample dilution: Pre-dilution and post-dilution

Internal bar code reader (optional)

Capable to communicate with LIS in bi-directional mode

Reaction System:

Cuvette: 40 disposable cuvettes

Reaction volume: 100~360µl

Operating temperature: 37 °C ±0.1 °C

ISE Module (optional)

Measuring K+, Na+, Cl-

Mixing Unit

Independent mixing bar

Optical System

Light Source: Halogen-tungsten lamp

Wavelength: 8 wavelengths, 340nm, 405nm, 450nm,

510nm、546nm、578nm、630nm、670nm

Absorption range: 0~4.0 Abs (10mm conversion), resolution

0.0001Abs

Stray Light 5.6Abs

Control and Calibration

Calibration modes: Linear (one point, two points and

multi-points), Logit-Log 4P, Logit-Log 5P, spline, exponential, polynomial, parabola

Control Rules: X-R, L-J, Westgard multi-rule, Cumulative sum

check, twin plot

Operation Unit

Operation system: Windows 8

Interface: RS-232

Working Conditions

Power Supply: 200~240V, 50/60Hz, ≤1000VA or 100~130V,

60Hz, ≤1000VA

Dimension: 690 mm (length) ×580 mm (depth) ×595 mm

(height)

Weight: 47 kg

Water Consumption: ≤ 2 L/H

Mindray Building, Keji 12th Road South, High-tech Industrial Park, Nanshan, Shenzhen 518057, P.R. China Tel: +86 755 8188 8998 Fax: +86 755 26582680

Mindray is listed on the NYSE under the symbol "MR"

E-mail: intl-market@mindray.com www.mindray.com

mindray
are registered trademarks or trademarks owned by Shenzhen Mindray Bio-medical Electronics Co., LTD.
@2015 Shenzhen Mindray Bio-Medical Electronics Co., Ltd. All rights reserved. Specifications subject to changes without prior notice.
P/N: ENG-BS-240-21285x8-20160105

